



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/802,906	03/17/2004	Gary K. Michelson	101.0058-03000	5309
22882	7590	07/30/2007		
MARTIN & FERRARO, LLP 1557 LAKE O'PINES STREET, NE HARTVILLE, OH 44632			EXAMINER SHAFFER, RICHARD R	
			ART UNIT 3733	PAPER NUMBER
			MAIL DATE 07/30/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/802,906

Applicant(s)

MICHELSON, GARY K.

Examiner

Richard R. Shaffer

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 30 April 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 7-41 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 7-41 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>5/16/2007</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 112

The amendments to the claims filed on April 30th, 2007 are acknowledged and accepted by the examiner. The previous rejections under 35 U.S.C. 112, second paragraph are hereby withdrawn.

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 38 and 39 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Claim 38 attempts to further limit claim 7 by stating that the bone-contacting surface is curved across the width and claim 39 states that the surface is at least in part flat across the width of the body. Since claim 7 allowed for two different conditions to meet the claim, it is not known which of the two or both are being modified by such statements. Further, if claim 39 was attempting to further limit the second condition, it fails to do so since claim 7 already states that a substantial portion is flat along the length and width.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

Art Unit: 3733

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 7-10, 13-16, 18-29, 35 and 37-41 are rejected under 35 U.S.C. 102(b) as being anticipated by Caesar (US Patent 2,825,329).

Caesar discloses a device (**Figures 1-3**) comprising: a low profile plate (**150**) with a convex surface (shown as opposite of the bone, but inherently capable of being flipped upside down as a matter of design choice) along the length (interpreted as being convex while looking down the length) of the plate; at least two bone screw receiving holes (**156**) receiving bone screws (**158**) arranged side-by-side; a locking element (**140**) capable of locking the screws (**158**) either through the use as shown in **Figures 1-3** or can also merely cover the top of plate (**150**) and be rotated, slid, etc into and out of alignment for allowing the screws through holes (**156**); locking element (**140**) having a circular head (**142**) with a cut-out (as seen in **Figure 2**).

Claims 7-29, 35 and 37-41 rejected under 35 U.S.C. 102(b) as being anticipated by Tornier (US Patent 4,488,543).

Tornier discloses a device (**Figures 1-3**) comprising: a low profile plate (**2**) with a bone contacting surface having a convex curvature along the length of the plate; **[First Interpretation]** three bone screw receiving holes (**6**) receiving bone screws (**3/3a**) arranged side-by-side; a locking element (**9** and **11** together); portion (**11**) of the locking element is elastic and thus can be pulled (partially rotated about an axis) out of the way of an inserting screw should one wish to do so; the locking element has a circular head (**11**) with a cutout (hole that allows **9** to pass through); **[Second Interpretation]** four bone screw receiving holes (**6** and **8**) receiving bone screws (**3/3a** and **9**); and a locking

Art Unit: 3733

element (11) which can be rotated in order to allow insertion of at least two bone screws (3/3a).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 30-34 and 36 are rejected under 35 U.S.C. 103(a) as being unpatentable over Caesar.

Caesar discloses all of the claimed limitations except for the use of a second locking element and thereby a second pair of bone screw receiving holes and making the device with a fusion promoting substance such as bone or bone morphogenetic protein.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to have provided for an additional locking element and plate should more than one fracture require fixation.

It would have further been obvious to one having ordinary skill in the art at the time the invention was made to use fusion promoting substances such as bone and bone morphogenetic protein (both well known in the art) to enhance bone healing.

Claims 30-34 and 36 are rejected under 35 U.S.C. 103(a) as being unpatentable over Tornier.

Tornier discloses all of the claimed limitations except for the use of a second locking element and thereby a second pair of bone screw receiving holes and making the device with a fusion promoting substance such as bone or bone morphogenetic protein.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to have provided for an additional locking element and plate should more than one fracture require fixation.

It would have further been obvious to one having ordinary skill in the art at the time the invention was made to use fusion promoting substances such as bone and bone morphogenetic protein (both well known in the art) to enhance bone healing.

Terminal Disclaimer

The terminal disclaimer filed on April 30th, 2007 disclaiming the terminal portion of any patent granted on this application, which would extend beyond the expiration date of Patent Numbers 7,118,573; 7,044,952; 7,041,105; and 6,527,776 has been reviewed and is accepted. The terminal disclaimer has been recorded.

Double Patenting

The following patents and patent applications have been considered for double patenting in the examination of this application:

Patents:

- 6,936,050 - 6,926,718 - 6,454,771 - 7,044,952 - 6,527,776 - 6,193,721
- 6,936,051 - 6,398,783 - 7,041,105 - 7,077,844 - 7,118,573 - 6,383,186
- 6,416,528 - 6,428,542 - 6,592,586 - 6,616,666 - 6,620,163 - 6,712,818

Art Unit: 3733

- 6,916,320 - 6,969,390 - 7,074,221 - 7,097,645 - 7,112,202 - 7,115,130
-6,139,550

Applications:

- 10/926,734 - 11/110,161 - 11/128,556

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 7-41 are rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over:

claims 1-173 of U.S. Patent No. **6,193,721**

claims 1-39 of U.S. Patent No. **6,936,051**

claims 1-117 of U.S. Patent No. **6,398,783**

in view of Reisberg (US Patent 5,468,242).

Comparison of Current Application against Conflicting Patents:

'721 Patent: Concave (in dependent claims, e.g. 7, 10), **Silent** in independent claims (1, 28) instead of **convex/flat**; locking element for covering first and second bone screw receiving holes.

'051 Patent: Concave instead of **convex/flat**; lock for at least two bone screw receiving holes.

'783 Patent: Concave instead of **convex/flat**; lock for at least two bone screws (inserted in two bone screw receiving holes).

Reisberg teaches (**Column 1, Lines 5-20**) that having implants that conform to the bone surface/anatomical shape (e.g. convex, flat, and concave) is desired. One having ordinary skill in the art at the time the invention was made would have developed a device that either was pre-shaped or moldable to conform to the anatomy to provide for a low profile fit that would 1) minimize external visualization of implant and therefore embarrassment of patient 2) irritation from inflammation 3) allow fixation members (screws, tacks, pins, etc) to have greater holding force (instead of having a significant portion hanging out of the bone). Therefore, applicant's numerous patents attempting to differentiate bone plates on the premise of concave vs. convex/flat are not novel. The mere shaping of an interface surface to conform to the intended bone to which it is to be placed is already known to be desired and beneficial.

Any additional difference between the previous patents and the current application lies in the fact that the claims (taking into account dependent claims in several) in the patents include more elements and are thus more specific. The inventions therefore define a "species" of the "generic" invention of claim 7. It has been

held that the generic invention is "anticipated" by the "species." See *In re Goodman*, 29 USPQ2d 2010 (Fed. Cir. 1993).

Claims 7-41 are provisionally rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over

claims 1-72 of copending Application No. **10/926,734**

claims 1-62 of copending Application No. **11/128,556**

in view of Reisberg. As discussed in the patent section, applicant is merely reciting a concave base section with one application being "at least one" again with the other being "at least two" in regard to the bone receiving holes. Each one recites at least "two bone screws" in regard to a locking element

This is a provisional obviousness-type double patenting rejection.

Response to Arguments

Applicant's arguments filed April 30th, 2007 have been fully considered but they are not persuasive. Applicant stated that Caesar failed to disclose a bone-contacting surface having a convex surface along the length or a flat surface along the length and width of the body. As explained in the current office action, there is a convex shaped surface to which it would be inherently capable of performing the claimed function. Applicant's other arguments are directed towards intended use arguments as well.

A recitation of the intended use of the claimed invention must result in a structural difference between the claimed invention and the prior art in order to patentably distinguish the claimed invention from the prior art. If the prior art structure is capable of performing the intended use, then it meets the claim.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Richard R. Shaffer whose telephone number is 571-272-8683. The examiner can normally be reached on Monday-Friday (7am-5pm).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Richard Shaffer
July 20th, 2007



EDUARDO C. ROBERT
SUPERVISORY PATENT EXAMINER